



TECHNICAL DATA SHEET

MICROSCOPE COVER LEICA

(EU) 2017/745 Annex XI-Part A Production Quality Assurance.

MICROSCOPE COVER LEICA

▶ Product Description	▶ Protective equipment used by doctors or nurses for surgical operations to prevent the possible risk of infection.
▶ Product Class	▶ (EU) 2017/745 Medical Device Regulation – Class Business Rule I
▶ Manufacturer's Location	▶ Tio Medikal 2/20 st. No.:53 (Begos 3. North Entrance, 35400 Buca OSB/Buca/İzmir)
▶ Purpose of usage	▶ It is used as a protective cover for microscopes to prevent contamination during surgical operations.

Quality

- Manufactured under ISO 13485:2016 and 13795-1 quality management standards.
- It has CE certificate.

Bio-Compatibility

- Does not contain latex.
- Sterilized with ethylene oxide.

Related Standard

- ISO 13485:2016 Medical Medical Devices Quality Management System
- 9001:2015 Quality Management System
- TS EN 13795 Surgical Clothing and Drapes
- TS EN ISO 11607-1:2017 Packaging for finally sterilised medical devices - Part 1: Rules for materials, sterile barrier systems and packaging systems / Products are made in accordance with the relevant standard.
- TS EN ISO 11607-2:2017 Packaging of medical devices-Finally sterilised- Part 2: Validation requirements for forming, sealing and joining

processes/ Products are made in accordance with the relevant standard.

Shelf Life

- 3 years

1. MICROSCOPE COVER LEICA SIZES




Figure 1. MICROSCOPE COVER LEICA Demonstration


















Available Materials: 04 Polyethylene 40 micron

	TYPE
REF . CODES	
222.10.000.01	68 mm

Tolerances: $\pm 3\%$ (ALL SIZES ARE PRODUCED & CUSTOMISATION)

Instruction for use:

1. The package of product shall be opened in sterile and aseptic conditions.
2. For a clean peel, open the package from the direction of the arrow slowly .
3. There are labels or marks on the drapes, which helps the user to follow the patient direction and unfolding directions.

	Do not use if package is damaged		Manufacturer		Single Use (Do not re-use)		Use by date
	Do not expose the product to sunlight.		Catalogue number		Consult instructions for use		Temperature Limitation
	Do not re-sterilize		Manufacturing Date		Sterilized using Ethyleneoxide and Single sterile barrier system		CE Marking
	Keep Dry		Batch Code		Caution		Medical device
	Unique device identifier						